



County of El Paso Purchasing Department
800 E. Overland Room 300
El Paso, Texas 79901
(915) 546-2048 / Fax: (915) 546-8180

ADDENDUM 1

To: All Interested Proposers

From: Linda Gonzalez, Inventory Bid Technician

Date: May 1, 2009

Subject: Bid # 09-033, Digital X-ray Machine for the Medical Examiner's Office

The Purchasing Department received a question relating to the above referenced proposal; the response to the following question:

1. The x-ray units listed for quoting are analog. These would require a CR System to create the digital image. Are there items missing on the bid specifications?

Yes. The specifications for the CR System were left out. See Attachment 1

2. The title of this bid is "Digital X-ray Machine for Medical Examiner's Office" However there is no mention in the model number or description from MinXray stating that the system is digital. Are you asking for digital images?

Yes. The model must be equivalent to DynaRad/Kodak .