



County of El Paso Purchasing Department
800 E. Overland Room 300
El Paso, Texas 79901
(915) 546-2048 / Fax: (915) 546-8180
www.epcounty.com

ADDENDUM 3

To: All Interested Vendors
From: Elvia Contreras, Formal Bid Buyer
Date: July 31, 2015
Subject: RFP# 15-030, Replacement of X-Ray Equipment & Related Software Systems for the Jail Annex

This addendum has been issued to notify vendors of the following questions and modifications:

- ❖ **The opening date has been extended to Tuesday, August 11, 2015 at 2:00 p.m.**
- ❖ **Deadline to questions has not been extended.**

1. Does the County know how many medical x-rays are done per year at the annex?

RESPONSE: Approximately 755.

2. The data cabling that will tie the different rooms together and the system to your network, will that be done by your maintenance people or by us? We know we are responsible for all interconnect cabling and cabling to a junction box.

RESPONSE: The vendor shall provide and install all data processing equipment and cables required for the operation of the new system.

3. Item 6. F: Are you requesting 1 digital panoramic system?

RESPONSE: Yes.

4. You are not requesting any digital intraoral system on bid?

RESPONSE: We do need a Digital Radiology machine.

5. Item 6. L: This equipment requires a digital network.

- A. Is this equipment tying into the existing network of the Annex or will it be on a separate network? We will need to get information from your IT personnel.

RESPONSE: The equipment will integrate into our private LAN, no separate network will be configured for this integration. Any requirements for connections outside of our private LAN will be the responsibility of the contractor to maintain and will require prior approval from the County Information Technology Department.

If on existing, are there ports available on existing router?

RESPONSE: Only four (4) network switch ports will be available for Private (County) LAN use, any additional will be the contractor's responsibility.

- B. Will the Annex's existing network support the installation of the new digital equipment and PACS, and be able to transfer large imaged data files?

RESPONSE: The County will support the delivery of large files inside our enterprise private LAN.

Will EPC install and support a new network (if required) to support newly installed equipment?

RESPONSE: No, the County will use existing enterprise network, if additional network equipment or cable drops are required it will be the vendors responsibility to provide and meet County specifications.

6. Item 7, second paragraph, we will need 100 amp 240vac single phase or 125 amp 208vac single phase for the x-ray generator.

RESPONSE: Yes we have the required power available.

7. Item 8, paragraph 3. The proposed equipment will have the same lead requirements as the present equipment. Your facility should have documentation on the rooms' compliance.

RESPONSE: There is documentation on tests done by the Radiation Safety tech (Gus Montes) for this area.

8. We need drawings showing dimensions of x-ray and dental area.

RESPONSE: Dimensions are as follows:

Dental Treatment Room: 14'x 20'

X-Ray Room 12'x 21'

Dark Room 6' x 9'

9. Item 6.d: For the 5 cassettes, what sizes do you want and how many of each?

RESPONSE: Six (6) cassettes total, 3 ea 14 x 17 & 3 ea 11 x 14 are needed.